



香港城市大學
City University of Hong Kong

專業 創新 胸懷全球
Professional · Creative
For The World

CityU Scholars

Self-sensing intelligent microrobots for noninvasive and wireless monitoring systems

Li, Zhongyi; Wang, Kun; Hou, Chaojian; Li, Chunyang; Zhang, Fanqing; Ren, Wu; Dong, Lixin; Zhao, Jing

Published in:

Microsystems and Nanoengineering

Published: 01/01/2023

Document Version:

Final Published version, also known as Publisher's PDF, Publisher's Final version or Version of Record

License:

CC BY

Publication record in CityU Scholars:

[Go to record](#)

Published version (DOI):

[10.1038/s41378-023-00574-4](https://doi.org/10.1038/s41378-023-00574-4)

Publication details:

Li, Z., Wang, K., Hou, C., Li, C., Zhang, F., Ren, W., Dong, L., & Zhao, J. (2023). Self-sensing intelligent microrobots for noninvasive and wireless monitoring systems. *Microsystems and Nanoengineering*, 9, Article 102. <https://doi.org/10.1038/s41378-023-00574-4>

Citing this paper

Please note that where the full-text provided on CityU Scholars is the Post-print version (also known as Accepted Author Manuscript, Peer-reviewed or Author Final version), it may differ from the Final Published version. When citing, ensure that you check and use the publisher's definitive version for pagination and other details.

General rights

Copyright for the publications made accessible via the CityU Scholars portal is retained by the author(s) and/or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights. Users may not further distribute the material or use it for any profit-making activity or commercial gain.

Publisher permission

Permission for previously published items are in accordance with publisher's copyright policies sourced from the SHERPA RoMEO database. Links to full text versions (either Published or Post-print) are only available if corresponding publishers allow open access.

Take down policy

Contact lbscholars@cityu.edu.hk if you believe that this document breaches copyright and provide us with details. We will remove access to the work immediately and investigate your claim.

ARTICLE

Open Access

Self-sensing intelligent microrobots for noninvasive and wireless monitoring systems

Zhongyi Li^{1,2}, Kun Wang³, Chaojian Hou³, Chunyang Li^{1,2}, Fanqing Zhang^{1,2}, Wu Ren⁴, Lixin Dong³✉ and Jing Zhao^{1,2}✉

Abstract

Microrobots have garnered tremendous attention due to their small size, flexible movement, and potential for various in situ treatments. However, functional modification of microrobots has become crucial for their interaction with the environment, except for precise motion control. Here, a novel artificial intelligence (AI) microrobot is designed that can respond to changes in the external environment without an onboard energy supply and transmit signals wirelessly in real time. The AI microrobot can cooperate with external electromagnetic imaging equipment and enhance the local radiofrequency (RF) magnetic field to achieve a large penetration sensing depth and a high spatial resolution. The working ranges are determined by the structure of the sensor circuit, and the corresponding enhancement effect can be modulated by the conductivity and permittivity of the surrounding environment, reaching ~560 times at most. Under the control of an external magnetic field, the magnetic tail can actuate the microrobotic agent to move accurately, with great potential to realize in situ monitoring in different places in the human body, almost noninvasively, especially around potential diseases, which is of great significance for early disease discovery and accurate diagnosis. In addition, the compatible fabrication process can produce swarms of functional microrobots. The findings highlight the feasibility of the self-sensing AI microrobots for the development of in situ diagnosis or even treatment according to sensing signals.

Introduction

Microrobots have attracted considerable attention for various biomedical applications in the near future. The small size and flexible movement characteristics urge microrobots to travel minimally invasively or essentially noninvasively in narrow and complex areas for in situ diagnosis and treatment, especially targeted drug delivery, which is difficult for existing equipment^{1–6}. Similar to the counterpart macro-robots, the major components for functionalizing a microrobot into a system still include perception, cognition, action, control, and more advanced intelligence^{7–18}. In previous studies, researchers have focused more on how to

realize precise motion control for microrobots, which relied significantly on global visual feedback and control from outside^{19–28}. Embedding self-perception into microrobots remains challenging, although it possesses unique value for local diagnosis for precision medicine. Considering the remarkably elevated concentration of biomarkers near the diseased tissue compared with diluted blood and other body fluids, self-sensing microrobots can be sent almost noninvasively to concerned positions. This can be beneficial for disease diagnosis, especially during the early stage. In addition to the fixed-point detection of the target position, microrobots can patrol the body to find abnormal situations in time. Furthermore, such functionalized microrobots can be used for in situ disease mapping, continuously monitoring the changes in a microenvironment, which is of great significance for the development of therapeutic approaches. Fluorescence sensing, an effective high-resolution sensing method, has been combined with microrobots for the detection of subcutaneous, intraocular, and superficial

Correspondence: Lixin Dong (lx.dong@cityu.edu.hk) or Jing Zhao (jingzhao@bit.edu.cn)

¹School of Mechatronical Engineering, Beijing Institute of Technology, 100081 Beijing, China

²Beijing Advanced Innovation Center for Intelligent Robots and Systems, Beijing Institute of Technology, 100081 Beijing, China

Full list of author information is available at the end of the article

These authors contributed equally: Zhongyi Li, Kun Wang, Chaojian Hou

© The Author(s) 2023



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

surfaces of some organs^{21,29,30}. However, single-photon excitation (ultraviolet range) can only provide a tissue interrogation depth of less than 0.5 mm; in contrast, multi-photon excitation (near-infrared range) can achieve a higher penetration rate, which is still less than 1.5–2 mm³¹. As a result, it is necessary to develop a new sensing method for microrobots to fulfill the requests for precise perception and real-time signal reading in vivo.

In this paper, we propose a wireless self-sensing artificial intelligence (AI) microrobot based on local magnetic field enhancement in electromagnetic imaging for noninvasive monitoring in vivo. Our functionalized microrobot consists of a head equipped with a sensor working as a tunable radiofrequency (RF) coil and a magnetic tail that can be driven by an external magnetic field. Considering the small size of the microrobot, the thin-film-electrode sensing device is manufactured as an inductor-capacitor circuit connected by through holes to enhance local signals in the electromagnetic imaging system and transmit signals wirelessly without supplying onboard power. Various sensor circuit structures have different working ranges according to the various electromagnetism characteristics, which can be affected by the surrounding environment. With increasing size and number of pairs of interdigital electrodes, the resonance frequency of the sensing circuit decreases. Moreover, the improved permittivity around the microrobot leads to an increase in the capacitance, causing a decrease in the resonant frequency of the sensor circuit. In contrast, an increase in environmental conductivity leads to less enhancement to the local RF field by increasing the resistance of the sensor circuit. Therefore, we can obtain changed environmental signals in real time to monitor various biochemical parameters. In particular, the variable pH and temperature of the solutions can effectively influence the resistance and resonance frequency of the sensor circuit, respectively, resulting in a higher conductivity of the acid solution with a lower pH value and a lower permittivity of the solution with a higher temperature. In addition, when the resonance frequency of the sensor is tuned to the operating frequency of the imaging device, bright spots appear in the position of the microrobot in the image, as shown in Fig. 1d. The novel self-sensing AI microrobot, with the potential to benefit from the advantages of existing electromagnetic imaging equipment in transmitting signals in the human body, provides a new route to realize passive wireless sensing in various complex positions, further simplifying the sensor structure and reducing the size.

Results and discussion

Figure 1a schematically shows the scenario of the AI microrobot patrolling inside the human body under external control. Among the existing electromagnetic imaging methods, magnetic resonance imaging (MRI) has been widely used because of its high penetration depth and spatial

resolution. During MRI imaging, the imaging signal strength is positively correlated with $\sin(\alpha)$, in which α is the flip angle of the excitation pulse. The strength of the electromagnetic field affects the angle during imaging and further determines the final imaging signal³². The sensing part at the head of the microrobot harvested the energy from the electromagnetic field generated by the RF coil of the imaging device and modulated the local RF magnetic field to change the signals at the corresponding position in the image^{33,34}. The dielectric characteristics of the surrounding environment, including the dielectric constant and conductivity, determine the electrical characteristics of the sensing circuit and further affect the enhancement effect on the local RF magnetic field. Diseases can change the dielectric properties of the micro-environment, and therefore, when the microrobot reaches the vicinity of the lesion, the enhanced signal can be changed, suggesting the occurrence of possible disease^{35,36}. Figure 1b shows a microrobot patterned on the substrate and an as-fabricated free-standing microrobot. The designed device is manufactured step by step following the processes shown in Fig. 1c. First, a 30 nm Al_2O_3 layer was deposited as the sacrificial layer on a preprepared Si substrate covered with 300 nm SiO_2 . Then, a double-layer tail containing a magnetic (Fe) and a prestress membrane (Si_xN_y) was deposited. Therein, the prestress Si_xN_y membrane was utilized as the main driving force of helical tail shape formation based on 3D rolling-up technology. 3D rolling-up technology is a reliable 3D forming technology resulting from the generation of the bending moment, along with the continuous release of the prestress Si_xN_y membrane. Here, a precompressive Si_xN_y membrane possessed a pair of opposite forces at the release interface, enabling an upward bending moment. With the isotropic release process of the initial parallelogram mesa, the helical shape can be successively formed. Owing to the final geometric structure morphology depending on the initial release shape design, the balance between the driving force and thin film stiffness, and so on, can be obtained^{37,38}. The head platform of the microrobot was made of 100 nm SiO_2 , which acted as the substrate to support the sensor. The three-layer structure of the sensor was sequentially fabricated on the head of the microrobot. The middle-interdigitated electrodes (<10 μm thickness) occupied approximately 25% of the head area and were surrounded by annular electrodes, which were connected to the bottom electrode through two via holes to form a closed resistance-inductance-capacitance (RLC) loop, as shown in Fig. 2a. The size of the RF coil ranged from ~500 to ~50 μm , and the number of interdigital electrode pairs could be adjusted according to the actual requirements, as shown in Supplementary Fig. S2. Finally, the chip was soaked in NaOH solution to etch the sacrificial Al_2O_3 layer, and the stress in the Si_xN_y layer was released, leading it to roll up into a helical tail. A more detailed fabrication process is shown in Supplementary Fig. S1.




Figure 2b demonstrates the principle diagram of the RLC passive sensor. L_s , represents the inductance introduced by the coil. R_s and C_s represent the sensing resistance and capacitance of the sensor introduced by interdigitated electrodes. The driving coil of the imaging equipment first generates the RF magnetic field, and the external loop coil of the sensor harvests the energy from the RF magnetic field and induces the current of the sensor circuit. The corresponding current can excite an additional magnetic field, which enhances the local RF magnetic field of the sensor. Considering that the distance between the excited coil and the sensor is far enough, the coupling coefficient is usually very small, resulting in the influence of the magnetic field generated by the sensor on the excited coil being negligible. Therefore, the excited magnetic field around the sensor, as well as the voltage generated by the sensor coil, can be considered a constant.

As a result, the current and magnetic field of the sensor merely depend on the impedance of the circuit, and the sensor circuit can obtain the maximum current and the most obvious enhancement effect on the local RF magnetic field at the resonance frequency. The resonance frequency (f_s) and quality factor (Q_s) can be calculated using the following equations:

$$f_s = \frac{1}{2\pi\sqrt{L_s C_s}}, Q_s = \frac{1}{R_s} \sqrt{\frac{L_s}{C_s}} \quad (1)$$

where L_s is a constant. C_s and R_s , introduced by interdigitated electrodes, can change with the surrounding environment, as shown in Fig. 2c (top). When the dielectric and conductive properties of the surrounding solution changed, the equivalent resistance and capacitance of interdigital electrodes responded accordingly. In

the equivalent circuit in Fig. 2c (bottom), R_l and C_l represent the resistance and capacitance when the electric field passes through the solution, respectively, while C_{sub} represents the substrate capacitance when the electric field passes through the substrate. The total capacitance C_{IDE} is the sum of C_l and C_{sub} . The equivalent impedance Z can be calculated using the following formulas:

$$C = C_{IDE} \tag{2}$$

$$R = R_l \tag{3}$$

$$Z = \frac{1}{j\omega C} // R = \frac{1}{j\omega C + 1/R} = \frac{1/R - j\omega C}{\omega^2 C^2 + 1/R^2} \tag{4}$$

and R_s and C_s in the sensor circuit can be calculated as follows:

$$R_s = \text{Re}(R_l // C_{IDE}) = \frac{1/R}{1/R^2 + \omega^2 C^2} = \frac{1}{1/R + \omega^2 C^2 R} \tag{5}$$

$$\omega C_s = -\frac{1}{\text{Im}(R_l // C_{IDE})} = \frac{1/R^2 + \omega^2 C^2}{\omega C} = \frac{1}{R^2 \omega C} + \omega C \tag{6}$$


in which R_s decreases with increasing ωC and ωC_s decreases with increasing R . R_s and ωC_s obtain the maximum and minimum values at $R = \frac{1}{\omega C}$ if ωC and R are

fixed, respectively. When $R > \frac{1}{\omega C}$, the increase in R leads to a decrease in R_s , and the increase in ωC leads to an increase in ωC_s . In contrast, when $R < \frac{1}{\omega C}$, the increase in R induces an increase in R_s , and the increase in ωC results in a decrease in ωC_s . Especially when $R \gg \frac{1}{\omega C}$:

$$R_s = \frac{1}{1/R + \omega^2 C^2 R} \approx \frac{1}{\omega^2 C^2 R} \tag{7}$$

$$\omega C_s = \frac{1}{R^2 \omega C} + \omega C \approx \omega C, C_s \approx C \tag{8}$$

According to the structural model, the electrical characteristics of interdigital electrodes were emulated by a high-frequency structure simulator (HFSS). The interdigital electrode model was extracted from the micro-robot with a head size of 500 μm and 11 interdigital electrode pairs. With the frequency varying from 0.5 to 2 GHz, the amplitudes of the real and imaginary components of the impedance decrease clearly, as shown in Supplementary Fig. S3. The increase in environmental conductivity will lead to an increase in the real impedance and a decrease in the imaginary impedance. Notably, when the conductivity increases from 0.001 to 0.1 S/m, the imaginary impedance is almost unchanged, which agrees with the theory of $R \gg \frac{1}{\omega C}$. The simulation results



are consistent with the theoretical analysis results at $R > \frac{1}{\omega C}$. Therefore, the increased environmental conductivity can lead the loop resistance to increase at the resonance frequency and weaken the enhancement effect.

The local electromagnetic field enhancement ability of the sandwich structure sensor, including the intermediate dielectric layer and top and bottom electrodes, can be simulated by the HFSS, as shown in Fig. 3a, in which an excited port is used to generate the RF magnetic field and the sensor surrounded by the dielectric as the environmental substitution is placed below the excited port. The substrate without the sensor is set for the control group, and deionized (DI) water ($\epsilon_r = 81, \sigma = 0\text{S/m}$) is used as the environmental dielectric. The local RF magnetic field strength can be evaluated by magnetic field strength integration. In Fig. 3b, the $500 \times 500 \mu\text{m}^2$ plane centered on the upper electrode is used to calculate the magnetic field strength. As the frequency increases, the strength integration of only the substrate model increases gradually. In contrast, some definite peaks appear for the sensor, illustrating the enhanced local RF magnetic field effect. According to the theoretical analysis, the greatest enhancement effect occurs at the resonance frequency, the first of which is ~ 1.24 GHz. The magnetic field strength integral for the substrate is $\sim 4.8\text{E-}7 \text{ T} \cdot \text{m}^2$ at 1.24 GHz, while the value can amplify ~ 45 times to $\sim 2.2\text{E-}5 \text{ T} \cdot \text{m}^2$ for the sensor. In addition, the enhancement amplitude reaches ~ 560 times at the second resonance point ($f = 4.94$ GHz). Moreover, the presence of the sensor greatly changed the local magnetic field distribution, especially in the area close to the sensor. The local RF magnetic field distribution of the sensor at 1.24 GHz is shown in Fig. 3c, d. The maximum electromagnetic field strength can reach ~ 300 A/m at 1.24 GHz. With increasing distance from the sensor, the corresponding strength decreases gradually. However, the maximum magnetic field strength is only ~ 3 A/m without the sensor, as shown in Fig. 3e, f. The magnetic field in the central area around the substrate is uniform and less than 1 A/m, while the magnetic field strength becomes larger at the boundary and the area closer to the excited port.

The magnetic field enhancement coefficient (MFEC), the ratio of the magnetic field strength integral with and without the sensor, can be utilized to evaluate the magnetic field enhancement effect at different planes, as shown in Eq. (9).

$$MFEC = \frac{\iint_D |H_{\text{sensor}}| dS}{\iint_D |H_{\text{substrate}}| dS} \quad (9)$$

The size of the plane is $500 \times 500 \mu\text{m}$, which is similar to the sensor at the center, and the height gradually increases from the upper substrate plane ($D = 0 \mu\text{m}$), where the MFEC is ~ 45.45 . As the height increases, the relevant

value of the MFEC decreases accordingly. When $D = 200 \mu\text{m}$, the MFEC changes to ~ 1.19 . When D increases further to $\sim 300 \mu\text{m}$, the MFEC decreases to ~ 1.03 , but the ability to change the magnetic field distribution remains strong according to the magnetic field distribution in Supplementary Fig. S4.

The sensor structure, including the size of the sensor and the number of interdigital electrode pairs, can affect the magnetic field strength, as shown in Supplementary Fig. S5. We compared the MFEC at $D = 0 \mu\text{m}$ with different sensor structures. The whole microrobot swam in the DI water. According to the simulation results, the resonance frequency peak increased rapidly from ~ 1 to ~ 30 GHz as the size of the sensor decreased from 500 to 50 μm . Moreover, the frequency peak has a left shift with increasing numbers of interdigital electrode pairs, which can promote the capacitance and reduce the resonance frequency. The statistical result of the shift in the resonance peak with the change in sensor structure is summarized in Supplementary Fig. S6. Overall, the small size of the sensor led to a high resonance frequency, resulting from the limitation of the placement space leading to smaller values of inductance and capacitance.

The surrounding environment can affect the local magnetic field enhancement performance of the sensor by changing the permittivity or conductivity. With the conductivity increasing exponentially from 0.001 to 1 S/m, under a fixed permittivity ($\epsilon_r = 81$), resonance peaks are concentrated at approximately 1.24 GHz when the swept frequency ranged from 0.1 to 2 GHz, as shown in Fig. 4a. The data at $D = 0 \mu\text{m}$ were used to calculate the MFEC. However, when the environmental conductivity is 10 S/m, the large resistance of the sensor circuit causes a small current, resulting in the local magnetic field remaining nearly unchanged. The relationship between the resonance frequency, the MFEC, and the conductivity is plotted in Fig. 4b based on the results from Fig. 4a. The MFEC curve was extracted at a frequency of 1.24 GHz (the red line in Fig. 4a, near the resonance peak). The stable resonance frequency with conductivity varying from 0.001 S/m to 0.1 S/m is mainly due to the small conductivity, where the solution impedance of the interdigital electrode R_l is much larger than $\frac{1}{\omega C_{IDE}}$. Correspondingly, the capacitance of the whole circuit C_s and the resonance frequency f_s remain unchanged. When the conductivity increased to 1 S/m, the corresponding resonance frequency decreased to 1.22 GHz. Moreover, the resonance frequency dropped rapidly to 0.24 GHz when the conductivity reached 10 S/m. Overall, the MFEC is more sensitive to small conductivity changes, demonstrating a monotonic downward trend with increasing conductivity at a frequency of ~ 1.24 GHz. It decreased from 103.6 at 0.001 S/m to 2.4 at 1 S/m but was only 0.98 at 10 S/m, indicating that there was no enhancement effect.




Fig. 3 HFSS simulation of the sensor. **a** Schematic diagram of the simulation model. The sensor is surrounded by the environmental dielectric, while the top exacted port generates the electromagnetic field. **b** The simulation results of the relationship between the local magnetic field intensity and frequency. The sensor-enhanced local magnetic field intensity will reach the maximum value at the resonance point. **c, d** The side and top views of the electromagnetic field distribution around the sensor at the first resonance point, respectively. **e, f** The contrast electromagnetic field distribution without a sensor at the first resonance point




Figure 4c shows the simulation results of permittivity from 1 to 81 under constant conductivity ($\sigma = 0$ S/m) with frequencies ranging from 0.1 to 4 GHz. With increasing permittivity, the resonance peak revealed a redshift from 3.7 to 1.24 GHz. Meanwhile, the MFEFC was obviously improved from 0.1 to over 100 at the resonance point. The detailed corresponding relationship between resonance frequency and permittivity is plotted in Fig. 4d. The decrease rate of the resonance frequency gradually slowed down with increasing permittivity. When the permittivity changed from 1 to 11, the resonance frequency dropped from 3.7 to 2.72 GHz. However, the frequency decreased by only 0.14 GHz (from 1.38 to 1.24 GHz) under permittivity varying from 71 to 81, suggesting that the tuned sensitivity of the resonance frequency provided high detection accuracy and response under conditions with low permittivity. To cooperate with the electromagnetic imaging system, we extracted MFEFC results with the

different dielectrics at 400 MHz, as shown in Fig. 4d (right), corresponding to MRI with operating magnetic field strength (B_0) of 9.4 T. Compared with the value near the resonance peak, the MFEFCs at 400 MHz are all less than 1, which means that the sensor weakens the local RF magnetic field instead, resulting from the induced magnetic field generated by the coil reducing the excited magnetic field. However, the induced magnetic field is large enough to amplify the field strength near the resonance peak. When the permittivity changes from 1 to 81, the MFEFC linearly increases from 0.11 to 0.96. A MFEFC of less than 1 can make the sensor appear as a dark spot in the electromagnetic imaging system while the surrounding environment presents as a bright background. Therefore, it is necessary to match the resonance peak with the working frequency of electromagnetic imaging to realize variation and easy observation for in vivo detection.




Fig. 5 Electromagnetic characteristics measurement of the sensor using the network analyzer. **a** The model of the PCB for the sensor measurement. A single-turn coil with a length of ~1 mm was connected to the network analyzer through the Cu electrode and SMA as the external coil. The sensor was immersed in different solutions dropped through the middle hole. **b** Optical images of the PCB for measurement. **c** Compared changes in the reflection coefficient S_{11} with and without the sensor when pH = 1. **d** S_{11} as a function of the frequency with the solution pH varying from 1 to 4. **e** Response characteristics of the sensor at different pH values. **f** Compared shifts of S_{11} with and without the sensor at 35 °C. **g** S_{11} as a function of the frequency with the water temperature varying from 35 to 55 °C. **h** Response characteristics of the sensor at different temperatures

On account of the theoretical results, we fabricated a suitable structure sensor on the microrobot and connected the device to the prepared external measuring equipment, as shown in Fig. 5a. A square coil with a side length of 1 mm was used as an external readout coil, which was coupled with the coil surrounding the perceptive interdigital electrodes. The external coil was connected to the network analyzer through a Sub-Miniature-A (SMA) connector for the frequency sweeping test, and the measured circuit was made on a printed circuit board (PCB). The chip carrying the device was fixed on the PCB by epoxy resin to prevent the solution from affecting the readout coil from the back. As shown in Fig. 5b, the solution can immerse the lower end of the PCB and contact the sensor through the front hole during the measurement. The reflection

coefficient S_{11} , which reflects the sensor response to the surrounding environment, can be calculated by Eq. (10):

$$S_{11} = 20 \log_{10} \left(\left| \frac{Z_{eq} - Z_0}{Z_{eq} + Z_0} \right| \right), Z_{eq} = R_r + j\omega L_r + \frac{\omega^2 M^2}{R_s + j\omega L_s + \frac{1}{j\omega C_s}} \quad (10)$$

Z_{eq} and Z_0 are the equivalent impedances of the readout coil and the system (normally 50 Ω), respectively. M is the coupling coefficient of the readout and the sensor coil. R_r and L_r are the resistance and inductance of the measured circuit, respectively. S_{11} reaches the minimum value near the resonance frequency when $Q_s \gg 1$ ³⁹. We prepared HCl solutions with different conductivities by changing the concentration, demonstrating the relevant pH value.

When the device was placed in the HCl solution at $\text{pH} = 1$ and $T = 25^\circ\text{C}$, the resonance frequency of the sensor centered at ~ 1.12 GHz, while the resonance frequency is ~ 1.85 GHz for the substrate only, as shown in Fig. 5c. As the pH increases from 1 to 4 in Fig. 5d, the curves shift to higher frequencies. The relationship between the resonance frequency and the pH is plotted in Fig. 5e based on the result from Fig. 5d. When the pH value decreases to 1, the resonance frequency rises to ~ 1.120 GHz. The free H^+ in the solution forms the diffusion capacitance, which attenuates the external electric field, resulting in the sensor capacitance decreasing instead of increasing⁴⁰. However, the resonance frequency remains nearly unchanged at ~ 1.110 GHz when the pH is close to 7 because of the low concentration of the solution, which leads to a larger resistance than the impedance generated by the capacitance. Nevertheless, the S_{11} at 1.110 GHz (the red line in Fig. 5d) as a function of pH in Fig. 5e monotonically increases from -3.35 dB to -3.17 dB when the pH increases from 1 to 4, indicating the sensitivity of the device as a pH sensor.

Moreover, the sensor can respond to ambient temperature changes. Solutions with different temperatures were obtained by heating the DI water, and the permittivity gradually decreased from ~ 80 at room temperature (25°C) to less than 70 at 60°C ^{41,42}. The resonance frequency moved from ~ 1.90 GHz to ~ 1.12 GHz when the device was fabricated on the substrate, which was surrounded by DI water at $T = 35^\circ\text{C}$, as shown in Fig. 5f. When the temperature increased from 35°C to 55°C , both the resonance frequency and the S_{11} of the device increased. Figure 5h demonstrates the relationship between the resonance frequency and the temperature derived from the measured result in Fig. 5g. The resonance frequency increases from 1.125 to 1.160 GHz linearly with increasing temperature due to the decreased permittivity, while S_{11} at 1.124 GHz monotonically changes from -7.99 to -6.11 dB, which is much larger than the change caused by the varying pH values.

Therefore, the sensor integrated into the microrobot can modulate the local magnetic field in the electromagnetic imaging process depending on the surrounding environment and affect the signals at the position in the image. Apparently, the AI microrobot can be guided to the prescribed destination and monitor the local micro-environment under the actuation magnetic field. We fabricated a series of microrobot integrated sensors with different sizes ($50\text{--}500\ \mu\text{m}$) and released them into DI water, as shown in Supplementary Fig. S8. Although the smaller microrobots could move in narrow channels and complex areas of the body, their resonance frequency was higher, causing difficulties while cooperating with the electromagnetic imaging equipment.

Locomotion control and posture adjustment of our integrated AI microrobots were carried out to attain the


motion characteristics of AI microrobots, as shown in Fig. 6 and Supplementary Movie 1. The actuation magnetic field was generated by three pairs of Helmholtz coils perpendicular to each other, and the intensity and direction of the resultant rotational field could be adjusted according to the motion requirements. The tail was rotated by the external rotating magnetic field, and the rotation was converted by the helical structure into a translational one so that the microrobot could move forward and backward according to the direction of rotation of the field. DI water was utilized as the work environment. During the whole driving experiment, both linear and circular locomotion was achieved from 0 s to 7 s and from 7 s to 23 s, respectively. In the first linear locomotion stage, a rotating uniform magnetic field in the yz plane was applied, enabling the AI microrobot to obviously spin and crawl like an active paddle accompanied by the external magnetic field stimulus. The magnetic field is represented as follows:

$$B_{yoz} = B_0 \sin(2\pi f_1 t) \vec{j} + B_0 \cos(2\pi f_1 t) \vec{k} \quad (11)$$

where B_0 is the magnetic induction intensity, $B_0 = 6$ mT; and f_1 is the rotation frequency, $f_1 = 1$ Hz. By comparing the position of the microrobot at 3 s and 7 s, the linear locomotion speed reached $133.32\ \mu\text{m/s}$. This large speed was mainly attributed to a tumbling motion based on a rotation axis at the interface line between the microsensor and substrate, leading to the $100\ \mu\text{m}$ length of the microsensor becoming the final rotation radius. Then, to achieve circular locomotion, a gradually increasing yaw angle with an increment speed of $24^\circ/\text{s}$ in the xy plane was applied to the initial B_{yoz} field. The total external magnetic field is as follows:

$$B_{\text{total}} = B_0 \sin(2\pi f_1 t) \sin(\omega t) \vec{i} + B_0 \sin(2\pi f_1 t) \cos(\omega t) \vec{j} + B_0 \cos(2\pi f_1 t) \vec{k} \quad (12)$$

where ω is the yaw angular speed. The results showed that a circular path started at 7 s and finished at approximately 22 s. Thus, the average angular speed can be calculated as $\omega = 2\pi/15 = 0.419$ rad/s. In addition, our integrated microrobot with a larger microsensor ($300\ \mu\text{m}$ length) was also successfully actuated by using the same external magnetic field, as shown in Supplementary Movie 2. The results showed that the optimized helical tails with larger geometry dimensions could offer a larger driving force to guarantee the tumbling motion of larger microsensors due to the larger magnetic material volume integrated into the helical tails. This phenomenon indicated that our manufacturing process was suitable for integrated microrobots with various scales. In conclusion, our integrated microrobot exhibited fast locomotion speed and flexible path planning. Next, the



formation mechanism of tumbling of our integrated AI microrobot was examined. In our integrated platform, a continued and stable driven moment, reliably formed as a result of the magnetic torque, always needed to align the magnetization of the microrobot with the applied field. However, the unsymmetrical geometric dimensions between the microsensors and helical tails lead to the position of the applied force always being located at the interfacial line between the microsensors and the Si substrate. With the help of the rotating magnetic field, the periodic friction force at the interface between the microsensors and the Si substrate enabled the AI microrobot to transform the rolling motion into linear motion in the plane, finally forming a paddling motion. This transformation motion characteristic is similar to the tumbling motion of peanut microrobots in a previous report²³. More importantly, due to the rotation of microsensors, the posture adjustment of our integrated AI microrobots can also be realized, enabling precise control of the relative rotation angle between the integrated coil and the external coil. This effective regulation behavior will promote the effective coupling between the coils and ensure the reliability of wireless signal extraction. Therefore, our integrated AI microrobots provide an effective integration strategy for future wireless monitoring systems with a controllable movement mode. In the future, optimized AI microrobot architectures need to be designed to achieve the conversion of helical rotational motion into linear motion, away from the substrate effect.

Our research strongly implied that the designed sensor integrated into the microrobot could realize controllable mobile sensing *in vivo* for disease diagnosis. The magnetic layer at the tail of the microrobot can generate force to actuate the microrobot under the excitation of an external magnetic field, and the wireless sensor can modulate the local RF magnetic field without providing external power. When the functional microrobot worked in the electromagnetic imaging system, the device position appeared as an artifact in the image, especially bright spots when the resonance frequency of the sensor matched the working frequency of the electromagnetic imaging. The conductivity and permittivity of the environment changed the modulation effect of the local RF magnetic field. Compared with *in vitro* detection, the abnormality around the lesion is more prominent, especially during the early stage of disease. Therefore, the designed device sent to the vicinity of the lesion can improve the detection accuracy of the disease. The AI microrobot is minute enough to enter the human body noninvasively and can realize detection in complex positions in the body. Owing to advanced micro/nanofabrication technology, large-scale parallel manufacturing can reduce cost and produce good consistency. In addition, the AI microrobot can cooperate with the electromagnetic imaging equipment to transmit the detection signal wirelessly, so the sensor itself only needs to locally interact with the RF magnetic field and modulate it. According to a previous study, the maximum working frequency of the MRI scanner for the human

body is ~ 500 MHz under an operating magnetic field strength (B_0) of 11.7 T⁴³. To effectively reduce the resonance frequency of the AI microrobot to match the MRI, we can increase the number of interdigital electrodes or increase the thickness of the upper electrode, as shown in Supplementary Fig. S7. Moreover, the refined structural design, including optimization of the inductance coil and loading of fixed capacitance and resistance, can facilitate cooperation with the MRI equipment.

Although the designed sensor was sensitive to the conductivity and permittivity change in the surrounding environment, which was usually related to the occurrence of diseases, it is not specific to different influence factors. Next, we will functionalize the upper electrode of the sensor by coating it with selective materials to specifically detect the concentrations of various ions, proteins, glucose, oxygen, and so on^{44–49}. After the functionalized sensor is combined with the marker, the permittivity and conductivity above the sensor will change. Therefore, the scene in which many microrobots, patrolling inside the body under external control, are navigated by MRI and actively find abnormal parts according to the sensing signal for in situ diagnosis and even treatment will be realized in the near future.

Conclusion

In summary, a wireless AI microrobot designed for noninvasive in vivo monitoring was fabricated to realize self-sensing functionality. The novel AI microrobot, consisting of a head sensor and a magnetic helical tail, can perceive environmental change sensitively based on the enhanced signal in electromagnetic imaging and can be precisely driven by an external actuating field. The sensor circuit harvests energy from the RF magnetic field without onboard power and is tuned or detuned by the conductivity and permittivity of the surrounding environment. This determines the enhancement effect on the local RF magnetic field, increasing it up to ~ 560 times. An increase in permittivity led to a corresponding decrease in the resonance frequency, while the increase in conductivity weakened the enhancement effect. Therefore, the impedance signal, affected by the surrounding environment, can reflect the working state of the microrobot in real time. Moreover, the magnetic helical tail actuated the microrobot to the target position under the control of an external magnetic field accordingly. Considering the AI microrobot's small size, controllable movement, and self-sensing properties, it has great potential to be used in noninvasive monitoring of the body combined with existing electromagnetic imaging equipment, especially those places that are difficult to reach by traditional equipment. Additionally, the integration of a passive sensor with the microrobot highlights the feasibility of noninvasive wireless sensing in situ, which is of great significance for the development of multiple therapeutic

means. Moreover, on account of the compatible fabrication process, mass-produced and cluster-controlled microrobots are predictable. Furthermore, integrating various functional coatings on the AI microrobot can enable the detection of various biochemical components and biomarkers for specific disease diagnoses in the future^{50,51}.

Methods

Device fabrication

The silicon substrate with a 300-nm-thick silicon dioxide film was first successively cleaned with acetone, isopropyl alcohol (IPA), and DI water. Then, a 30-nm Al_2O_3 film was deposited on the substrate using e-beam evaporation as a sacrificial layer. A 30-nm silicon nitride (Si_xN_y) membrane with compressive stress was grown with plasma-enhanced chemical vapor deposition (PECVD). Then, the magnetic helical tails were made of a Fe membrane by using UV lithography, development, metallization by magnetic sputtering and lift-off. To achieve helical tails with different dimensions, the balance between the Fe membrane and the prestrained Si_xN_y needed to be optimized. Here, we used a 150-nm-thick Fe film at a sputtering deposition rate of 0.15 nm/s. Then, inductively coupled plasma (ICP) dry etching was utilized to etch the Si_xN_y membrane until the Al_2O_3 layer was exposed, obtaining batched helical tail mesa arrays. The etching gas was CHF_3 , and the etching power was 200 W.

A 100 nm SiO_2 layer was deposited by electron beam evaporation, and the lower electrode of the sensor (100 nm Au) was fabricated via UV lithography, development, thermal evaporation, and lift-off. Another photoresist was then patterned as a soft mask for SiO_2 etching. The SiO_2 layer was etched by ICP to form the head platform of the microrobot. Then, another 100 nm SiO_2 dielectric layer was deposited by electron beam evaporation and patterned by photolithography, etching, and photoresist removal processes to serve as the intermediate dielectric layer for the sensor. The second SiO_2 layer was etched by reactive ion etching (RIE). The etching gas was CF_4 , and the power was 150 W. The upper patterned electrodes (300 nm Au) were deposited by thermal evaporation after UV lithography.

Finally, the sample was immersed in a 45% sodium hydroxide solution to etch the Al_2O_3 layer. Both the microsensors and helical tails were separated from the Si substrate after 2 h, making the microrobots free-standing in the solution. Finally, we used DI water to clean the microrobots and transferred them to the swimming pool by the injector for the following magnetic driving experiment.

Device simulation

The electromagnetic simulation software HFSS was used to simulate the device's performance. The device was placed in the horizontal center of the simulation

environment, and the excited port for generating the RF magnetic field was located above at least 6 times the device width to avoid the influence coming from the device being excited. In addition, the horizontal dimension was at least 4 times the device width so that the surrounding boundary was far enough away from the local area to avoid affecting the local RF magnetic field uniformity of the device. The simulation structure only included the three-layer structure of the sensor. To facilitate the calculation of the magnetic field integration, we directly defined the calculation plane with different D values in the model to obtain the magnetic field distribution, whose size was consistent with the calculation plane. The integral of the magnetic field can be calculated by the HFSS, and simulation models of sensors with different structures were created. The corresponding response of the sensor to various conductivities and permittivities was simulated by changing the environmental dielectric properties.

Device measurement

The PCB was connected to the network analyzer through an SMA connector. Considering the small size of the sensor, it was in direct contact with the PCB, leading to a close range between the sensor and the readout coil of $\sim 10\ \mu\text{m}$. The covered epoxy resin not only fixed the chip on the PCB but also played a sealing role in protecting the solution from contacting the PCB from the back, resulting in parasitic capacitance. The hole in the center of the PCB can allow the sensor to soak in solution. The clean substrate without any structures was used as the control group. Solutions with different pH values were obtained by gradually diluting the purchased 0.1 mol/L HCl solution, and the corresponding pH value was confirmed by a pH meter. We used the arc funnel to inject the solution from the top, which can be discharged from the end under the immersed PCB. In contrast, to test the influence of different temperatures, the end of the PCB was placed in a beaker filled with DI water, whose temperature can be tuned by water-bath heating. The network analyzer was used for the frequency sweep under different environmental conditions.

Acknowledgements

This work was supported by the National Key R&D Program of China (Grant No. 2019YFB1309701), the National Natural Science Foundation of China (NSFC, Grant Nos. 3020021501839 and 61804009), and the Beijing Institute of Technology Research Fund Program for Young Scholars, Beijing Institute of Technology Analysis & Testing Center, and SUSTech Core Research Facilities.

Author details

¹School of Mechatronical Engineering, Beijing Institute of Technology, 100081 Beijing, China. ²Beijing Advanced Innovation Center for Intelligent Robots and Systems, Beijing Institute of Technology, 100081 Beijing, China. ³Department of Biomedical Engineering, City University of Hong Kong, 999077 Kowloon Tong, Hong Kong, China. ⁴School of Integrated Circuits and Electronics, Beijing Institute of Technology, 100081 Beijing, China

Author contributions

Z.L. designed the experiment, fabricated and characterized the sensors, and wrote the manuscript. K.W. and C.H. designed and fabricated AI microrobots, characterized their motion abilities and wrote the manuscript. C.L. and F.Z. helped with the fabrication and characterization of the sensors. W.R. helped with the characterization of the sensors. L.D. and J.Z. supervised the research, designed the experiment, and reviewed the manuscript.

Conflict of interest

The authors declare no competing interests.

Supplementary information The online version contains supplementary material available at <https://doi.org/10.1038/s41378-023-00574-4>.

Received: 31 March 2023 Revised: 1 June 2023 Accepted: 10 July 2023
Published online: 09 August 2023

References

- Chang, D. et al. Biologically targeted magnetic hyperthermia: potential and limitations. *Front. Pharmacol.* **9**, 831 (2018).
- Gultepe, E. et al. Biopsy with thermally-responsive untethered microtools. *Adv. Mater.* **25**, 514–519 (2013).
- Wu, Z. et al. A swarm of slippery micropropellers penetrates the vitreous body of the eye. *Sci. Adv.* **4**, eaat4388 (2018).
- Abdelmohsen, L. K. E. A., Peng, F., Tu, Y. & Wilson, D. A. Micro- and nanomotors for biomedical applications. *J. Mater. Chem. B* **2**, 2395–2408 (2014).
- Zheng, Z. et al. Ionic shape-morphing microrobotic end-effectors for environmentally adaptive targeting, releasing, and sampling. *Nat. Commun.* **12**, 411 (2021).
- Luo, M., Feng, Y., Wang, T. & Guan, J. Micro-/nanorobots at work in active drug delivery. *Adv. Funct. Mater.* **28**, 1706100 (2018).
- Wang, B., Kostarelos, K., Nelson, B. J. & Zhang, L. Trends in micro-/nanorobotics: materials development, actuation, localization, and system integration for biomedical applications. *Adv. Mater.* **33**, 2002047 (2021).
- Li, Z., Li, C., Dong, L. & Zhao, J. A review of microrobot's system: towards system integration for autonomous actuation in vivo. *Micromachines* **12**, 1249 (2021).
- Avci, E., Grammatikopoulou, M. & Yang, G.-Z. Laser-printing and 3D optical-control of untethered microrobots. *Adv. Opt. Mater.* **5**, 1700031 (2017).
- Li, D., Liu, C., Yang, Y., Wang, L. & Shen, Y. Micro-rocket robot with all-optic actuating and tracking in blood. *Light Sci. Appl.* **9**, 84 (2020).
- Oh, J., Park, J.-O., Park, S. & Ko, S. Y. Image-based guidance system for intravascular microrobot: fiducial marker-based registration using biplanar fluoroscopic images & CTA images. In *Proc. 2015 15th International Conference on Control, Automation and Systems (ICCAS)*, 919–922. <https://doi.org/10.1109/ICCAS.2015.7364753> (IEEE, 2015).
- Yamanaka, T. & Arai, F. Self-propelled swimming microrobot using electro-osmotic propulsion and biofuel cell. *IEEE Robot. Autom. Lett.* **3**, 1787–1792 (2018).
- Kim, D. et al. Propulsion and rotation of microrobot based on a force on a magnetic material in a time-varying magnetic field using a wireless power transfer system. *IEEE Trans. Magn.* **56**, 1–5 (2020).
- Miskin, M. Z. et al. Electronically integrated, mass-manufactured, microscopic robots. *Nature* **584**, 557–561 (2020).
- Reynolds, M. F. et al. Microscopic robots with onboard digital control. *Sci. Robot.* **7**, eabq2296 (2022).
- Li, J. et al. Biohybrid micro- and nanorobots for intelligent drug delivery. *Cyborg Bionic Syst.* **2022**, 2022/9824057 (2022).
- Wang, H., Kan, J., Zhang, X., Gu, C. & Yang, Z. Pt/CNT micro-nanorobots driven by glucose catalytic decomposition. *Cyborg Bionic Syst.* **2021**, 2021/9876064 (2021).
- Yu, Z. et al. Memensors with a crossbar structure for in-vivo H_2O_2 detection. In *Proc. 2022 IEEE International Conference on Manipulation, Manufacturing and Measurement on the Nanoscale (3M-NANO)*, 334–339. <https://doi.org/10.1109/3M-NANO56083.2022.9941702> (IEEE, 2022).

19. Wang, B. et al. Endoscopy-assisted magnetic navigation of biohybrid soft microrobots with rapid endoluminal delivery and imaging. *Sci. Robot.* **6**, eabd2813 (2021).
20. Aziz, A. et al. Real-time IR tracking of single reflective micromotors through scattering tissues. *Adv. Funct. Mater.* **29**, 1905272 (2019).
21. Xie, L. et al. Photoacoustic imaging-trackable magnetic microswimmers for pathogenic bacterial infection treatment. *ACS Nano* **14**, 2880–2893 (2020).
22. Aziz, A. et al. Medical imaging of microrobots: toward in vivo applications. *ACS Nano* **14**, 10865–10893 (2020).
23. Xie, H. et al. Reconfigurable magnetic microrobot swarm: multimode transformation, locomotion, and manipulation. *Sci. Robot.* **4**, eaav8006 (2019).
24. Pané, S. et al. Imaging technologies for biomedical micro- and nanoswimmers. *Adv. Mater. Technol.* **4**, 1800575 (2019).
25. Zhang, Y. et al. Real-time tracking of fluorescent magnetic spore-based microrobots for remote detection of *C. diff* toxins. *Sci. Adv.* **5**, eaau9650 (2019).
26. Wang, L. et al. Reconfigurable vortex-like paramagnetic nanoparticle swarm with upstream motility and high body-length ratio velocity. *Research* **6**, 0088 (2023).
27. Wang, L. et al. Micro-nanocarriers based drug delivery technology for blood-brain barrier crossing and brain tumor targeting therapy. *Small* **18**, 2203678 (2022).
28. Dai, Y. et al. Magnetically actuated cell-robot system: precise control, manipulation, and multimode conversion. *Small* **18**, 2105414 (2022).
29. Esteban-Fernández De Ávila, B. et al. Single cell real-time miRNAs sensing based on nanomotors. *ACS Nano* **9**, 6756–6764 (2015).
30. Ergeneman, O. et al. In vitro oxygen sensing using intraocular microrobots. *IEEE Trans. Biomed. Eng.* **59**, 3104–3109 (2012).
31. Marcu, L. Fluorescence lifetime techniques in medical applications. *Ann. Biomed. Eng.* **40**, 304–331 (2012).
32. Hai, A., Spanoudaki, V. C., Bartelle, B. B. & Jasanoff, A. Wireless resonant circuits for the minimally invasive sensing of biophysical processes in magnetic resonance imaging. *Nat. Biomed. Eng.* **3**, 69–78 (2019).
33. Chen, J. C. et al. A wireless millimetric magnetoelectric implant for the endovascular stimulation of peripheral nerves. *Nat. Biomed. Eng.* **6**, 706–716 (2022).
34. Chen, P.-J., Saati, S., Varna, R., Humayun, M. S. & Tai, Y.-C. Wireless intraocular pressure sensing using microfabricated minimally invasive flexible-coiled LC sensor implant. *J. Microelectromech. Syst.* **19**, 721–734 (2010).
35. Cheng, Y. & Fu, M. Dielectric properties for differentiating normal and malignant thyroid tissues. *Med Sci. Monit.* **24**, 1276–1281 (2018).
36. Sugitani, T., Arihiro, K. & Kikkawa, T. Comparative study on dielectric constants and conductivities of invasive ductal carcinoma tissues. In *Proc. 2015 37th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)* 4387–4390. <https://doi.org/10.1109/EMBC.2015.7319367> (IEEE, 2015).
37. Yu, Z. et al. Modeling and characterization of artificial bacteria flagella with micro-structured soft-magnetic teeth. In *Proc. 2022 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS)*, 1062–1067. <https://doi.org/10.1109/IROS47612.2022.9981435> (IEEE, 2022).
38. Wang, K. et al. 3D chiral micro-pinwheels based on rolling-up kirigami technology. *Small Methods* **7**, 2201627. <https://doi.org/10.1002/smt.202201627> (2023).
39. Zhu, H.-T., Chen, Y., Xiong, Y.-F., Xu, F. & Lu, Y.-Q. A flexible wireless dielectric sensor for noninvasive fluid monitoring. *Sensors* **20**, 174 (2019).
40. Shamsul Arefin, M. et al. A MEMS capacitive pH sensor for high acidic and basic solutions. In *Proc. 2014 IEEE SENSORS*, 1792–1794. <https://doi.org/10.1109/ICSENS.2014.6985373> (IEEE, 2014).
41. Jacobsen, R. E., Arslanagić, S. & Lavrinenko, A. V. Water-based devices for advanced control of electromagnetic waves. *Appl. Phys. Rev.* **8**, 041304 (2021).
42. Catenaccio, A., Daruich, Y. & Magallanes, C. Temperature dependence of the permittivity of water. *Chem. Phys. Lett.* **367**, 669–671 (2003).
43. Le Bihan, D. & Schild, T. Human brain MRI at 500 MHz, scientific perspectives and technological challenges. *Supercond. Sci. Technol.* **30**, 033003 (2017).
44. Fang, Y. et al. Fabrication of a needle microsensor and its applications in the detection of dissolved oxygen. *J. Sens.* **2015**, 1–7 (2015).
45. Mou, L., Xia, Y. & Jiang, X. Epidermal sensor for potentiometric analysis of metabolite and electrolyte. *Anal. Chem.* **93**, 11525–11531 (2021).
46. Cao, Y. et al. Binding-regulated click ligation for selective detection of proteins. *Biosens. Bioelectron.* **78**, 100–105 (2016).
47. Chen, C. & Lehr, J. Label-free selective detection of protein markers in the picomolar range via a convenient voltammetric sensing strategy. *Electroanalysis* **33**, 563–567 (2021).
48. Gao, W. et al. Fully integrated wearable sensor arrays for multiplexed in situ perspiration analysis. *Nature* **529**, 509–514 (2016).
49. Khatayevich, D. et al. Selective detection of target proteins by peptide-enabled graphene biosensor. *Small* **10**, 1505–1513 (2014).
50. Vaisocherová, H. et al. Ultralow fouling and functionalizable surface chemistry based on a zwitterionic polymer enabling sensitive and specific protein detection in undiluted blood plasma. *Anal. Chem.* **80**, 7894–7901 (2008).
51. Jang, H.-J. et al. Electronic cortisol detection using an antibody-embedded polymer coupled to a field-effect transistor. *ACS Appl. Mater. Interfaces* **10**, 16233–16237 (2018).